SPECIFICATIONS

Reciprocating Holder Apparatus 7 – 12x12

The 12-row reciprocating holder system is a modified Apparatus 7 instrument ideal for testing modified release-rate dosage forms including medical devices and transdermal systems.

12x12 Specifications

Altitude	0 - 2000 m (0 - 6562 ft.)
Temperature	5 to 40 °C
Humidity (non-condensing)	Not more than 80 % RH
Voltage requirements	115 V / 60 Hz 230 V / 50 Hz
Current requirements	115 V—2.0 Amp 230 V—2.0 Amp
Fuse requirements	115 V—2.0 Amp, 250 V, 5 mm x 20 mm FAST 230 V—2.0 Amp, 250 V, 5 mm x 20 mm FAST
Minimum dips per minute	5
Maximum dips per minute	60
Stroke length	2 ± 0.1 cm
Water bath temperature	Ambient + 5 to 55 °C
Number of rows	12
Number of sample positions	12
Run time display format	hhh:mm programmable up to 999:59
Dip time interval display format	hhh:mm programmable up to 999:59
Drain time display format	mm:ss programmable up to 99:59
Hold dip time display format	mm:ss programmable up to 99:59
Printer	Impact
Dimensions	Height: 73.66 cm (29 in.) Width: 68.58 cm (27 in.) Depth: 80.65 cm (31.75 in.)
Equipment weight	44.5 kg (98.1 lbs)

To learn more about Dissolution Testing,
visit us at www.agilent.com/lifesciences/dissolution

This information is subject to change without notice.
© Agilent Technologies, Inc. 2011
Printed in U.S.A., August 31, 2011
5990-7410EN

The Measure of Confidence

Agilent Technologies